

M93U Ultrasound puncture pad II

CVC Insertion Simulator II-U



Caution

Don't mark on the model and other components with pen or leave printed materials contacted on their surface.

Ink marks on the models will be irremovable.

Instruction Manual



Table of contents

Before you start

Manufacturer's note	P.1
Set includes	P.2
Dos and Dont's	P.3

[I . Ultrasound puncture pad II]

Introductory ultrasound training block	
“REAL VESSELS”	P.5
Parts names and functions	P.6
Preparation	P.7-P9
After a session	P.10-P12
Storage and maintenance	P.12

[II . Transparent cannulation block]

Parts names and functions	P.14
Preparation	P.15-P16
After a session	P.17

Troubleshooting	P.18
-----------------------	------

Manufacturer's

The simulator is designed for training in CVC procedures.

Any other use, or any use not in accordance with the enclosed instructions, is strongly discouraged. Kyoto Kagaku Co., Ltd. cannot be held responsible for any accident or damage resulting from such use. Please use this model carefully and refrain from subjecting it to any unnecessary stress or wear.

Should you have any questions or concerns regarding use of this simulator, please contact the distributor you purchased from or Kyoto Kagaku Co., Ltd.

• Features

The simulator comes with 2 kinds of training pads for relevant area and an introductory ultrasound training block.

■ Ultrasound puncture pad II

Both internal jugular and axillary veins are accessible under ultrasound scanning. Anatomically correct vein and artery relationship.
Veins collapse under light pressure of the probe.

■ Transparent cannulation block


Facilitates three-dimensional anatomical understanding.
An effective training tool for developing guide wire insertion skills.

■ Introductory ultrasound training block REAL VESSELS

An introductory training block to acquire basics of ultrasound-guided puncture.

Set includes

Before your first use, ensure that you have all components listed below.



A. male upper torso manikin.....	1	G. 50ml syringe.....	1
B. transparent anatomical block.....	1		
C. skin for cannulation training	1		
D. ultrasound puncture pad II	2		
E. introductory ultrasound training block	1		
“REAL VESSELS”			
F. irrigation bottle	1		

Before you start

DOs and DON'Ts

 DOs and DON'Ts	
DOs Use new needles for training. Recommended needle size: 23G or thinner Durable transparent cannulation pad is highly recommended for insertion training in order to maintain the puncture pad longer. Handle with care. The materials for the simulator are a special composition of soft resin. Handle them with the utmost care at all times. Clean the surface of the pad with dry or wet soft cloth and put talcum powder before storage. Clean the upper torso manikin by dry cloth. Store the training set at room temperature, away from heat, moisture and direct sunlight.	DON'Ts Do not peel the skin from the puncture pad. Never wipe the simulator with thinner or other organic solvent. Don't mark on the simulator and the components with pen or leave any printed materials in contact with their surface. Ink marks on them will not be removable. Do not store the simulator and pads with any water left inside. Do not use broken or bent needles for training.
 Caution When you use the puncture pads; after aspirating do not return the fluid from syringe into simulator. It may damage the vessel tube and may lead to leakage.	

Ultrasound puncture pads II



Introductory ultrasound training block “REAL VESSELS”

- 1. General information P. 5
- 2. Preparation P. 5
- 3. Training and after a session P. 5

Ultrasound puncture pad II

Parts names and functions P. 6

Preparation

- 1. Place the pad into the torso P. 7
- 2. Fill the artery tube with water P. 7-P. 8
- 3. Fill the vein tube with water P. 8-P. 9

After a session

- 1. Discharge water from the artery tube P. 10
- 2. Discharge water from the vein tube P. 11
- 3. Detach the puncture pad from the simulator P. 12
- 4. Maintenance of the skin of the pad P. 12

Find the how-to-use video on our website

<http://www.kyotokagaku.com>

REAL VESSELS

Introductory ultrasound training block “REAL VESSELS”

1 General


This block facilitates training in basics of ultrasound guided punctures, before moving onto trainings with the anatomical type ultrasound pad.

Features

- 2 simulated vessel lines: straight and curving.
- Both lines are embedded with slope to represent from shallow vessel to deep one.
- Vessel wall yields under pressure of a needle tip.

Training items


- How to get correct ultrasound images. Probe manipulation.
- Basics for ultrasound-guided vessel access.



2 Preparation




Peel the protective sheet carefully.




Pour water into the wider slit in the container and fill it up to the line on the wall. Ensure that shallower end of vessels are fully under the water surface.

3 Training and after a session



Spread some ultrasound gel on the surface of the block and start training. When the needle tip is in the vessel, your syringe can collect water. Add water to the container slit as necessary.




After use, wash off the gel completely with running water and dry well. Then, replace the protective sheet before storage.



.....
Do not leave paper, cloth or other materials except for the included protective sheet contact with the block surface. Such items may stick to the surface and lead to damage the block.

Ultrasound puncture pad II



Caution

.....
Vein tube embedded in the pad has a soft and thin wall to realize life-like ultrasound image.

Do not push the probe strongly.

Make sure that the tubes are not folded or tucked between walls of the pad and the body torso. If fluid or air is injected while the tube is folded or clogged, excessive pressure may cause a breakage or burst of the thin tube.


Fill the vessels with COLORLESS water.

After aspirating, do not return fluid from syringe into simulator.


Preparation

Place the pad into the torso Fill the artery tube with water

1 Place the pad into the torso



1. Put the tubes through the hole of the head side.




2. Open the cut of left shoulder and put through the tubes. Place the pad into the torso.



.....
Make sure that the tubes are not folded or tucked between walls of the pad and the body torso. The simulator doesn't work properly when the tubes are folded.


2 First, fill the artery tube with water



1. Open the tube clamp at the head end of the artery tube.



2. Fill a container or a jar with 200cc of water.



Open the clamp



.....
Put the torso upright. This is to avoid bubbles being caught in the tube as well as to prevent the water flows out from its open end.



Preparation

Fill the artery tube with water
Fill the vein tube with water


2 First, fill the artery tube with water



2. Connect the water-filled 50ml syringe to the connector at the lower end of the artery tube (upper and rear side tube at the shoulder). Screw in the tip of the syringe to the connector. Advance the piston slowly until the water surface reaches height of the tube clamp.




3. When the tube has been filled by water, close the tube clamp.
Take off the syringe from the tube.



3 Then, Fill the vein tube with water



1. Put some water (around 5cm height) in the irrigation bottle.
Connect the tip of the tube from the bottle to the head end of the vein tube.




2. Connect the syringe filled with water to the connector at the lower end of the vein tube. Screw in the tip of the syringe clockwise to the connector until it locks.

3 Then, Fill the vein tube with water



3. Push the piston of the syringe slowly until the water flows out into the bottle. Take off the syringe from the tube. Keep the bottle connected while the simulator




4. Lay the simulator and start your training session.




.....
When the vessel tube is empty, ultrasound image will not be shown properly.




1 Discharge water from the artery tube




1. Put the torso upright and connect the empty 50ml syringe to the lower end of the artery tube.



2. Open the tube clamp.



3. Drain water by pulling the piston of the syringe. Detach the syringe from the tube and discharge the collected water.




.....
Ensure that all the water is discharged from the pad after use.
Do not store the pad while any water remains inside as it may lead to malfunction or damage the pad.


2 Discharge water from the vein tube




1. Put the torso upright and disconnect the tip of the tube from the irrigation bottle. Connect the empty 50ml syringe to the lower end of the vein tube. Drain water by pulling with the syringe.



2. Detach the syringe and discharge the water.




3. Tilt the body torso toward the syringe to aspirate remaining water in the tubings.




.....

Ensure that all the water is discharged from the pad after use.
Do not store the pad while any water remains inside as it may lead to malfunction or damage the pad.

3 Detach the puncture pad from the simulator



1. Open the cut of left shoulder and remove the tubes. Pull out the puncture pad from inferior side.




2. Remove the puncture pad from the torso.



Do not lift the pad by pulling the vessel tubes as this may cause breakage in the embedded tube walls.

4 Maintenance of the skin of the pad



1. To remove the ultrasound gel on the skin, wipe the skin surface gently by soft cloth, or soft wet close. If stickiness persists, wipe with soft cloth with small amount of alcohol for disinfection.



2. The skin of the ultrasound pad is not fixed and may come off if you hold the pad by skin only. In such case, replace the skin not to leave air bubbles under it. Air bubble or layer between skin and soft tissue will impair the quality of scanning image.



Never use thinner or organic solvent for cleaning. It may cause breakage in the simulator skin. Do not touch the surface of the skinless pad, or expose the surface to air for a long time. Dust attached to the surface may weaken its adhesion and lead to deterioration in scanning images.

Transparent cannulation block





Parts names and		P.14
functions General		P.15
information		
Preparation		P.15
Remove the pad currently set on the torso		P.16
Set up the cannulation block		P.17
Detach the pad from the torso		

Find the how-to-use video on our website

<http://www.kyotokagaku.com>

Transparent cannulation block



General information

The transparent cannulation block is an effective educational model to facilitate understanding of the relevant anatomical structures and is applicable to trainings in cannulation by interchanging with the puncture pads.

Training items

- Anatomical understanding
- Learn the appropriate depth and angle of the needle for each approach.
- Handling and manipulation of the catheter (guide wire).
- Simulate the steps of the procedures.



- Do not make any puncture at any site besides the prepared openings.
- The transparent block is not designed to be filled with fluid.
- Do not pour any fluid or water to the openings on the cannulation block.
- Please handle the skin sheet with utmost care. Excessive strain may cause breakage.

1 Remove the pad currently set on the torso


If other puncture pad is already set on the torso, take it out following the instruction on:

Remove the ultrasound puncture pad




P.12

2 Set up the cannulation block




Place the transparent cannulation block to the cavity.




Now you can practice cannulation by observing the catheter through the transparent block.

Cover the pad with the skin sheet



Cover the cannulation block with the skin sheet. Put the skin sheet over the pad so that the slits of the skin sheet fit with the openings on the pad.




Now, the simulator is ready for cannulation training. The slits in the skin sheet allow trainees to remove it with the catheter inserted in the pad.




Caution

.....
**Do not make any puncture at any site besides the prepared openings.
The transparent block is not designed to be filled with fluid.
Do not pour any fluid or water to the openings on the cannulation block.
Please handle the skin sheet with utmost care.
Excessive strain may cause breakage.**

3 Detach the pad from the torso



1. Remove all needles and catheters from the simulator.
Remove the skin.



2. Remove the pad holding its frame.

- If you are moving onto next training session using other kind of training pad, set up the new pad following the instruction on:

Set up the ultrasound puncture pad



P. 7

Trouble shooting

Quick check-up before calling the customer service. Use the table if you have problems using the simulator. Look in this section for a description of the problem to find a possible solution. (TEL +81-75-605-2510)



Trouble	Possible Reason	What to Do
I cannot fill/discharge the fluid to/from the vessel tubes properly.	The body torso with the pad is laid down.	Put the torso upright.
	One or some vessel tubes from the pad are folded.	Straighten the tubes.
	The puncture pad is worn out.	Order new pads for replacement.
Heavy leakage from puncture area.	The puncture pad is worn out.	Order new pads for replacement.
	Pool of fluid is formed in the pad.	Straighten the tubes. Withdraw all fluid from the tubes and squeeze the remaining fluid from the soft tissue. Dry the pad naturally.
Ultrasound image is unclear.	The vessel tubes are not filled by water/fluid. Or, air bubbles are formed in the vessels.	Fully fill the vessels with water.
	Air layer is formed under the skin sheet.	Reattach the skin sheet removing air bubbled and layers.
	The puncture pad is worn out. Order new pads for replacement.	Order new pads for replacement.



Caution

Ink from this manual can stain the model - do not let it touch the model and its components.

For inquiries and service, please contact your distributor or KYOTO KAGAKU CO., LTD.

Consumables and replacement parts

Part number	Description
11347-170	Ultrasound puncture pad II (a set of 2)
11347-190	Transparent cannulation block
11347-210	Introductory ultrasound training block (a set of 2)
11347-320	Irrigation bottle for ultrasound puncture pad



WEB: www.kyotokagaku.com e-mail: rw-kyoto@kyotokagaku.co.jp

Worldwide Inquiries & Ordering

Kyotokagaku Head Office and Factories:
15 Kitanechoya-cho, Fushimi-ku, Kyoto, 612-8388, JAPAN
Tel: +81-75-605-2510 Fax: +81-75-605-2519

All America regions: USA, Canada, and South America



WEB: www.kkamerica-inc.com E-MAIL: info@kkamerica-inc.com
3109 Lomita Boulevard, Torrance, CA 90505-5108, USA
Tel: 1-310-325-8860 Fax: 1-310-325-8867
(Toll free in North America: 877-648-8195)

Europe, Russia & Africa

Kyoto Kagaku Europe GmbH.

TEL: +49-69-5060-28160
De-Saint-Exupery-Str.10, 60549 Frankfurt, Germany

The contents of the instruction manual are subject to change without prior notice.

No part of this instruction manual may be reproduced or transmitted in any form without permission from the manufacturer.

Please contact manufacturer for extra copies of this manual which may contain important updates and revisions.

Please contact manufacturer with any discrepancies, typos, or mistakes in this manual or product feedback. Your cooperation is greatly appreciated.

July.2026